

THE STATE TRADING CORPORATION OF INDIA LIMITED
(A Government of India Enterprise)
5th Floor, Mogul's Court, Basheerbagh,
Hyderabad-500 001.

Tender Notice No: STC/HYD/GID/2018-19/04
Dated 20.08.2018

Corrigendum No. 1 dated: 31.08.2018

In the Tender for Supply of Forensic and other Scientific Equipment to AP Police, the following amendments are made:

In the list of items tendered on page 1, the following amendments to EMD are made:

Item No.2. FTIR Spectrophotometer with Library – EMD - Rs.55,000/-

Item No.3. U V - VIS Spectrophotometer – EMD – Rs. 78,000/-

Item No.44. FT-IR spectrophotometer – EMD – Rs. 2,20,000/-

Clause 12 (A): documents are required to be uploaded online as part of Technical Bid Part

Point 2 may be read as

“Scanned Copy of Pay Order / Demand Draft/Bid Bond for the amounts given in the list of items on page 1 of the tender document towards Earnest Money Deposit or Copy of MSME/NSIC Certificate mentioning the name of the item quoted. Name & Number of the Item to be clearly mentioned. Original DD should be submitted at STC, Hyderabad Office before date of close of tender i.e, **05.09.2018**”.

Other Items & Conditions

Clause (C): Earnest Money Deposit, may be read as

“Earnest Money Deposit (EMD) as per the amounts given in the list of items on page 1 of the tender document in the Form of Pay Order / Demand Draft/ Bid Bond in favour of “The STC of India Ltd., Hyderabad”. However, for sake of convenience the foreign suppliers may submit the requisite EMD through their Indian Agents in equivalent Indian Currency, should they so desire. The EMD in the form of DD/Payorder/Bid Bond to be placed in a sealed envelope along with a covering letter duly superscribing the Tender No. Item & Name of Item addressed to The Branch Manager, STC of India Ltd., Mogul's Court, 5th Floor, Basheerbagh, Hyderabad 500 001 and dropped in the Tender Box placed conspicuously on the 5th floor, Mogul's Court, Basheerbagh, Hyderabad 500 001.

50. SPECIFICATIONS FOR STEREO ZOOM MICROSCOPE WITH DIGITAL CAMERA

S. NO.	SPECIFICATIONS FOR OPTICAL SYSTEM	
1	Type	Greenough
2	Zoom ratio	4.4:1 or better
3	Eyepieces	10x
4	Viewing angle	60° or better
5	Magnification range	8x to 35x or better

6	Objective	1x
7	Resolution	Minimum 150 Lp/mm or better
8	Working distance	90mm or higher
9	Field of view	Minimum range from 5.8mm to 25mm diameter or better
10	Illumination	LED lights for Transmitted and Reflected light sources with auto shutoff

CAMERA SPECIFICATIONS

S.NO.	SPECIFICATIONS	
1	Resolution	1920(v)X1080(h) or better pixel
2	Type	Digital 5 Megapixel or more
3	Sensor	CMOS
4	Interfaces	Should have SD card, USB, Ethernet, HDMI.
5	Software	Imaging Software for basic image capture and simple measuring and annotation applications.

Note:-Vendors with above specifications or better can participate.

Warranty: 3 years comprehensive warranty from the date of installation i.e. spares, service, supported software update and calibration/validation etc.

Training: Hands on trained should be given to the staff members during installation at the customer site.

Service support: The maintenance service of the instrument /software should be done by qualified & trained service engineers and breakdown of the instrument/software should be attended within 48 hours of registering the complaint.

Vendor should also ensure spares and service support, at least for 5 years after the completion of the warranty period.

Note: Compliance Statement: The vendor should give a compliance statement for each specification and should also enclose relevant supporting literature in the form of company broacher/printed matter highlighting & marking the technical specifications.

The vendor should be ready to provide demo as & when required by the technical committee at the time of technical evaluation.

The camera and Microscope should be of reputed company.
